




Canine IL-2 ELISA Reagent Set

NB-06-1141

FOR RESEARCH USE ONLY

Technical Notes:	This kit is for the quantitative measurement of Canine IL-2 in cell culture supernatants. If assaying other sample types, an appropriate Sample and Standard Diluent will need to be developed and validated. Any changes to the ELISA protocol may significantly affect the results generated and will require optimization.		
Included Components:	Description	Quantity	Component Number
	Canine IL-2 Coated Plate	2 each	VS0258D-CP
	Canine IL-2 Standard	2 each	VS0258D-ST
	Canine IL-2 Detection Antibody	2 each	VS0258D-DA
	Streptavidin-HRP	1 each	AR0068-001
	Plate Sealer	6 each	N/A
Additional Reagents Required:	Reagent	Formulation	
	DPBS	0.008M sodium phosphate, 0.002M potassium phosphate, 0.14M sodium chloride, 0.01M potassium chloride, pH 7.4	
	Standard and Sample Diluent	Complete cell culture medium used to generate cell culture supernatant samples. It is critical that this medium contain at least 1% carrier protein. If the medium does not contain carrier protein, use Reagent Diluent to dilute the Standard and samples.	
	Reagent Diluent	4% BSA in DPBS, 0.2 µm filtered	
	Wash Buffer	0.05% Tween®-20 in DPBS	
	Substrate	3,3',5,5'-tetramethylbenzidine (TMB) Substrate	
	Stop Solution	0.18 M Sulfuric Acid	
Component Preparation:	Component	Preparation	
	Canine IL-2 Standard	Reconstitute Standard in 1 mL Standard and Sample Diluent. Dilute 125 µL of the reconstituted standard in 375 µL of Standard and Sample Diluent. The Standard now has a concentration of 800 pg/mL . Prepare 1:1 serial dilutions of the Standard by mixing 250 µL Standard with 250 µL Standard and Sample Diluent. Repeat 1:1 serial dilutions until reach a final concentration of 12.5 pg/mL. Use Standard and Sample Diluent as a zero standard.	
	Canine IL-2 Detection Antibody Working Solution	Reconstitute Detection Antibody in 500 µL Reagent Diluent. Dilute the 500 µL of reconstituted Detection Antibody in 11.5 mL Reagent Diluent.	
	Streptavidin-HRP Working Solution	Dilute 500 µL of Streptavidin-HRP in 11.5 mL Reagent Diluent.	

<p>ELISA Procedure:</p>	<ol style="list-style-type: none"> 1. Prepare Standard and cell culture supernatant sample dilutions in Standard and Sample Diluent. 2. Add 100 uL of Standard or sample to appropriate wells. <p>Note: Run each Standard or sample in duplicate.</p> <ol style="list-style-type: none"> 3. Cover plate with Plate Sealer and incubate at room temperature (20-25C) for 1 hour. 4. Wash plate FOUR times with Wash Buffer. <p>Note: Gently squeeze the long sides of plate frame before washing to ensure all strips remain securely in the frame. Empty plate contents. Use a squirt wash bottle to vigorously fill each well completely with 1X Wash Buffer, then empty plate contents. Repeat procedure three additional times for a total of FOUR washes. Blot plate onto paper towels or other absorbent material.</p> <ol style="list-style-type: none"> 5. Add 100 uL of Detection Antibody Working Solution to each well. 6. Cover plate with Plate Sealer and incubate at room temperature for 1 hour. 7. Wash plate FOUR times with Wash Buffer as described in step 4. 8. Add 100 uL of Streptavidin-HPR Working Solution to each well. 9. Cover plate with Plate Sealer and incubate at room temperature for 30 minutes. 10. Wash plate FOUR times with Wash Buffer as described in step 4. 11. Add 100 uL of TMB Substrate Solution to each well. 12. Develop the plate in the dark at room temperature for 30 minutes. <p>Note: Do NOT cover plate with Plate Sealer.</p> <ol style="list-style-type: none"> 13. Stop reaction by adding 100 uL of Stop Solution to each well. 14. Measure absorbance on a plate reader at 450 nm. 													
<p>Typical Standard Curve:</p>	 <p style="text-align: center;">Typical Standard Curve</p> <table border="1"> <caption>Data points for Typical Standard Curve</caption> <thead> <tr> <th>Canine IL-2 (pg/mL)</th> <th>Average OD (450 nm)</th> </tr> </thead> <tbody> <tr><td>0</td><td>0.0</td></tr> <tr><td>100</td><td>0.2</td></tr> <tr><td>200</td><td>0.4</td></tr> <tr><td>400</td><td>0.8</td></tr> <tr><td>800</td><td>2.1</td></tr> </tbody> </table>	Canine IL-2 (pg/mL)	Average OD (450 nm)	0	0.0	100	0.2	200	0.4	400	0.8	800	2.1	<p>Data represents a typical standard curve generated using the Canine IL-2 ELISA Development Kit.</p> <p>A standard curve should be generated with each assay.</p>
Canine IL-2 (pg/mL)	Average OD (450 nm)													
0	0.0													
100	0.2													
200	0.4													
400	0.8													
800	2.1													
<p>Representative Data:</p>	<p style="text-align: center;">Stimulant Canine IL-2 (pg/mL) Unstimulated <12.5 Staphylococcal enterotoxin B (SEB; 5 µg/mL) 259 Phytohemagglutinin (PHA; 10 µg/mL) <12.5 Phorbol 12-myristate 13-acetate (PMA; 10 ng/mL) and Ionomycin (500 ng/mL) 5991</p>	<p>PBMCs harvested by ficoll density gradient from an apparently healthy canine were suspended in RPMI medium containing 10% serum and stimulated as desired. The cell-free supernatants were harvested following three days stimulation and analyzed in the Canine IL-2 ELISA Development Kit.</p>												
<p>Country of Origin:</p>	<p>USA</p>													